



Flange mounted transducer "FMT", maximum torque: 10 Nm



Order data

Order number	654350 10
GTIN	2050001016676
Item class	66D

Description

Version:

- **Static sensors complete with cable (2 m) for direct connection to No. 654300 or No. 654340 (automatic detection of calibration data).**
- **Direct fixing to the workbench with 3 screws (not included).**

Supplied with:

UKAS calibration certificate on the right.

Up to and including size 150 with screw tightening simulator for checking electric and pneumatic screwdrivers. In a sturdy plastic case.

Note:

Not suitable for impact wrenches.

Test certificate: Manufacturer's test certificate

Calibration: O4

Measuring range: 0.5 - 10 Nm

Square socket: 1/4 inch

Hexagon socket: 1/4 inch

Technical description

Square socket	1/4 inch
maximum torque	10 Nm
Measuring range	0.5 - 10 Nm
Hexagon socket	1/4 inch
Calibration	O4

Direction of tightening	Right-hand tightening
Torque measuring accuracy	± 0.5 %
Test certificate	Manufacturer's test certificate

Services

DAkkS calibration Torque analyser / sensor maximum torque 1000/2 Nm	018840 1000/2
Calibration Torque analyser / sensor maximum torque 400 Nm	018860 400
DAkkS calibration Torque analyser / sensor maximum torque 1000 Nm	018840 1000
DAkkS calibration, right Torque analyser / sensor Type 1000	020130 1000
DAkkS calibration on both sides Torque analyser / sensor maximum torque 1000/2 Nm	020150 1000/2
Calibration, right Torque analyser / sensor Type 400	020120 400

Accessories

Screw tightening simulator maximum torque 10 Nm	654352 10
---	-----------